

Dräger Alcotest 6810 med

Measuring accuracy

The measuring accuracy of the Dräger Alcotest 6810 is at least:

Breath alcohol concentration:

Reproducibility with an ethanol standard:

0 to 0,50 mg/l:

0,008 mg/l

≥ 0,50 mg/l:

1,7 % of measurement value