

March 30, 2020

Supplement to the letter SARS-CoV-2 and handling of medical devices for Intensive Care of 24 March 2020

Dear Sir or Madam,

We would like to draw your attention to the fact that the recommendations in section A1 of the above-mentioned letter are product-specific and do not refer to the clinical condition or treatment methods of the patient.

It remains the responsibility of the clinical institution to evaluate the therapeutic measures and act accordingly. Furthermore, we would like to remind you that increased attention must be paid for sufficient humidification performance of intubated patients. This could be achieved for example by active humidification or a Heat-and-Moisture-Exchanger with integrated Filter (HMEF TwinStar®).

We kindly ask you to take this information into consideration and follow the respective Instructions for Use.

If you have any further questions or need any support, please contact Dräger's local service support or sales representative.

Kind Regards,



Meike Lessau
Global Product Manager
Business Unit Hospital
Consumables & Accessories
Medical Division



Melanie Ruge
Global Business Manager
Business Unit Hospital
Consumables & Accessories
Medical Division