



DEPARTMENT OF HEALTH & HUMAN SERVICES

NIOSH Reference: TN-20767
Mfr. Reference: DRA213

Centers for Disease Control
and Prevention (CDC)

National Institute for Occupational
Safety and Health (NIOSH)
National Personal Protective
Technology Laboratory (NPPTL)
626 Cochran Mill Road
Pittsburgh, PA 15236-0070
Phone: 412-386-4000
Fax: 412-386-4051
March 7, 2017

Mr. Klaus-Michael Rück
Research & Development of PPE
Dräger Safety AG & Co. KGaA
Revalstraße 1
D-23560 Lübeck
GERMANY

Dear Mr. Rück:

The National Institute for Occupational Safety and Health (NIOSH) has reviewed your request accepted March 25, 2016. This request was for approval of the model X-plore 8000 Powered Air Purifying Respirator (PAPR) system with the Dräger FPS 7000 series full facepieces (part numbers R56502, R56310 and R56503) with chemical cartridge (part number 6739590) for protections against chlorine, hydrogen fluoride, ammonia, methylamine, formaldehyde, hydrogen sulfide, sulfur dioxide and hydrogen chloride (CL/HF/AM/MA/FM/HS/SD/HC). Reference assembly matrix DRA216AMd.xls dated February 17, 2017.

Testing was successfully performed under TN-20766.

This request is granted. Approvals are granted only for documentation written in the English language. It is the manufacturer's responsibility to correctly translate materials desired in languages other than English. Approval number TC-23C-3220 has been assigned. This respirator is approved for protection against chlorine (CL), hydrogen fluoride (HF), ammonia (AM), methylamine (MA), formaldehyde (FM), hydrogen sulfide (HS), sulfur dioxide (SD) and hydrogen chloride (HC), (CL/HF/AM/MA/FM/HS/SD/HC).

The final respirator and cartridge approval labels are included as attachments under TN-20770. The cautions and limitations which apply to these approvals are on the approval labels. Only those assemblies affected by this request, or where new approval numbers are assigned, apply to this approval action. Production approval labels cannot include information on unapproved configurations.

The approved assemblies consist of the parts as listed on the approval labels and the assembly matrix. Parts are to be marked with the numbers indicated on the approval label in a legible and permanent manner (marking cannot be removed without evidence of its previous presence).

This certificate of approval is not an endorsement of the respirator by NIOSH, and such endorsement shall not be stated or implied in advertisements or other publicity. However, you may publicize the fact that this respirator has met the requirements of Title 42, *Code of Federal Regulations*, Part 84 (42 CFR 84).

No changes may be made to any respirators and accompanying documentation without prior written approval of NIOSH. Requests for changes must be submitted to NIOSH and a modification of this approval must be granted before changes are made.

Sincerely,

A handwritten signature in cursive script, appearing to read "David Chirdon".

David Chirdon
Chief, Conformity Verification and
Standards Development Branch
National Personal Protective Technology Laboratory