

Certificate of Compliance

Certificate Number **20111207-E180059**
Report Reference **E180059-20110217**
Issue Date **2011-December-07**

Page 1 of 2



Issued to: **DRAEGER SAFETY AG & CO KGAA**
REVALSTRASSE 1
23560 LUEBECK GERMANY

*This is to certify that
representative samples of*

GAS AND VAPOR DETECTION EQUIPMENT CLASSIFIED FOR USE IN HAZARDOUS LOCATIONS


Gas detection transmitters, Models ETR 02xx, ETR 03xx, ITR 02xx, ITR 03xx, ITR 04xx, ITR 05xx, XTR 02xx, and XTR 03xx, XTR 04xx, XTR 05xx, where xx stand for different models (instead of xx, it may also be used an asterisk **) for use in Class I, Groups A, B, C, and D; Class II, Groups E, F, and G; Class III Hazardous Locations.


*Have been investigated by Underwriters Laboratories in accordance with
the Standard(s) indicated on this Certificate.*

Standard(s) for Safety: **Explosion-Proof and Dust-Ignition-Proof Electrical Equipment for Use in
Hazardous (Classified) Locations - UL 1203**

Additional Information: See UL On-line Certification Directory at WWW.UL.COM for additional information.

Only those products bearing the UL Classification Mark for the U.S. and Canada should be considered as being covered by UL's Classification and Follow-Up Service and meeting the appropriate U.S. and Canadian requirements.

The UL Classification Mark includes: the UL in a circle symbol:  with the word "CLASSIFIED" (as shown); a control number (may be alphanumeric) assigned by UL; a statement to indicate the extent of UL's evaluation of the product; and the product category name (product identity) as indicated in the appropriate UL Directory. The UL

Classification Mark for Canada includes: the UL Classification Mark for Canada:  with the word "CLASSIFIED" (as shown); a control number (may be alphanumeric) assigned by UL; a statement to indicate the extent of UL's evaluation of the product; and the product category name (product identity) in English, French, or English/French as indicated in the appropriate UL Directory.

Look for the UL Classification Mark on the product

William R. Carney

Director, North American Certification Programs

Underwriters Laboratories Inc.

Any information and documentation involving UL Mark services are provided on behalf of Underwriters Laboratories Inc. (UL) or any authorized licensee of UL.

For questions, please contact a local UL Customer Service Representative at <http://www.ul.com/global/eng/pages/corporate/contactus>

Certificate of Compliance

Certificate Number **20111207-E180059**
Report Reference **E180059-20110217**
Issue Date **2011-December-07**

Page 2 of 2



This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.

Transmitter Junction Box, Models EAC 01, followed by 0 or 1, followed by x, where x stand for different models (instead of x, it may also be used an asterisk *) for use in Class I, Groups A, B, C, and D; Class II, Groups E, F, and G; Class III Hazardous Locations.

William R. Carney
Director, North American Certification Programs

Underwriters Laboratories Inc.

Any information and documentation involving UL Mark services are provided on behalf of Underwriters Laboratories Inc. (UL) or any authorized licensee of UL.

For questions, please contact a local UL Customer Service Representative at <http://www.ul.com/global/eng/pages/corporate/contactus>