

## Resuscitaire® Radiant Warmer

### Free Standing and Wall Mount configurations

The Dräger Resuscitaire® warmer combines neonatal thermoregulation and respiratory support into a single, versatile device for Labor & Delivery procedures. The ergonomic design puts all the components you need for clinical emergencies and resuscitation at your fingertips – helping you support newborns at a highly critical time.



Free Standing Mobile

## Benefits

---

### Choice of configurations

In addition to its original design, the Resuscitaire is also available in Free Standing Mobile and Wall Mounted (does not include a resuscitation package) versions to meet the needs of your hospital. A variety of options and accessories lets you further configure the unit to your specific requirements.

---

### Progressive respiratory support

The device's progressive respiratory support capabilities help you stabilize patient breathing while providing the minimal amount of support required to safeguard fragile lungs. The unit provides T-Piece resuscitation – the new standard for neonatal resuscitation. The AutoBreath option allows you to perform resuscitation hands-free: set the breath per minute, FIO<sub>2</sub>, PEEP and airway pressure relief. When the AutoBreath interface is comfortably secured this will provide hands free ventilatory support to your baby. All these parameters are adjustable based on the clinical needs of the infant.

---

### Fast, uniform thermoregulation

The Resuscitaire provides fast heating due to its unique silica quartz technology – a highly-responsive heating element. A full-length parabolic reflector helps distribute the heat evenly over the entire mattress area. A pre-warm mode lets you focus on the mother and baby, rather than having to silence alarms.

## Benefits (continued)

### Ergonomic design optimizes workflow

**Free Standing:** The blender, suction, flowmeter, and positive pressure are all mounted on the stand and located at your fingertips. Front-facing gauges accurately monitor airway pressure, suction, and gas supply to help you control the process. The suction compartment is located at the back of the unit to make routine cleaning quick and easy. This transportable model can be used to warm both mom and baby after birth and support kangaroo care.

**Wall Mount:** This easy-to-use, compact design option fits into any birthing room design that does not require resuscitation.

## Standard Features

### Free Standing Mobile

- Height 74"
- Warmer Head Rotation +/- 90°
- Apgar Timer
- Prewarm Mode
- Manual Mode
- Baby/Servo Mode
- Integrated Exam Light

### Options:

- Basic Resuscitation
  - Includes Flowmeters (2)
  - Suction
  - Manometer
  - Tank Holders
  - Integrated Regulators (2)
- Integrated Blender
- T-piece Resuscitation
- IV Pole
- Shelf

### Wall Mounted Warmer

- Warmer Head Rotation +/- 90°
- Apgar Timer
- Prewarm Mode
- Manual Mode
- Baby/Servo Mode
- Integrated Exam Light



MT-1814-2005

**TECHNICAL SPECIFICATIONS DRÄGER RESUSCITAIRE WARMER – Free Standing with (or without) Resuscitation**
**PHYSICAL ATTRIBUTES (without options/accessories)**

Height	74 in (188 cm)
Width	28 in (71.1 cm)
Depth	46 in (116.8 cm)
Weight (without options/accessories)	259 lbs (117.9 kg)

**MAXIMUM POWER REQUIREMENTS**

Model 120V	120V, 50 – 60 Hz, 750 W
220/240V	220 – 240V, 50 – 60 Hz, 750 W

**LED EXAMINATION LIGHT**

Illumination	>0.11 lumens/cm <sup>2</sup> (100 ft. Candles)
--------------	--

**CONTROLLER HEATING ALARMS**

Check patient	System alerts for 5 minutes in Manual Mode every 30 seconds after 10 minutes
Baby temperature	±1° C from Set Point
High temperature	Skin Temp. 39.0 ±0.2 °C
Probe	Short or open circuit/No probe
System fail	Indicates System Fail
Power fail	AC Power Interruption

**ALARM SILENCE/RESET INTERVALS**

Check patient	Resets clock for 10 minutes manual mode
Baby temperature	10 minutes
High temperature	2 minutes
Procedural silence	Presilences baby temp alarm for 5 minutes

**SET TEMPERATURE/HEATER OUTPUT**

Set temp range	34 – 38 °C (>37.0 °C override)
Heater output	0 – 100%, 10% increments
Skin temp display	18 – 43 °C
Accuracy	±0.2 °C (31 to 37 °C)

**RESUSCITATION MODULE**

Wall supply pressure	40 – 75 psi (275 – 517 kPa)
Cylinder pressure	2,900 psi max (19,994 kPa)

**PATIENT GAS SUPPLY**

Flow control range	0 – 15 LPM
Airway pressure limit	
Adjustable	0 – 50 cm H <sub>2</sub> O (0 – 4.9 kPa)
Internal preset limit	60 cm H <sub>2</sub> O (5.9 kPa)

**AUXILIARY FLOW CIRCUIT**

Auxiliary flow range	0 – 15 LPM
Fixed internal	160 cm H <sub>2</sub> O (15.7 kPa)

**SUCTION CIRCUIT**

Adjustable	0 – 150 mmHg (0 – 20 kPa)
Disposable collection container	40 ml

**ALERTS**

Manual mode	System alerts for 5 minutes every 30 seconds after 10 minutes
Apgar timer	Alerts at 1, 5 and 10 minutes
Timer	Visual timer up to 1 hour

**IV POLE**

Maximum load	5 lbs (2.2 kg)
--------------	----------------

**TECHNICAL SPECIFICATIONS DRÄGER RESUSCITAIRE WARMER – Wall Mount****SPECIFICATIONS**

Length	29.8 in (75.7 cm)
Width	10 in (25.4 cm)
Height	6 in (15.2 cm)
Weight (without options/accessories)	<35 lbs (15.92 kg)

**MAXIMUM POWER REQUIREMENTS**

Model 120V	120V, 50 – 60 Hz, 750 W
220/240V	220 – 240V, 50 – 60 Hz, 750 W

**LED EXAMINATION LIGHT**

Illumination	> 100 Ft. Candles (0.11 Lumens/cm <sup>2</sup> )
--------------	--

**CONTROLLER HEATING ALARMS**

Check patient	System alerts for 5 minutes in Manual Mode every 30 seconds after 10 minutes
Baby temperature	±1° C from Set Point
High temperature	Skin Temp. 39.0 ±0.2 °C
Probe	Short or open circuit/No probe
System fail	Indicates System Fail
Power fail	AC Power Interruption

**ALARM SILENCE/RESET INTERVALS (Manual Mode)**

Check patient	Resets clock for 10 minutes
Baby temperature	10 minutes
High temperature	2 minutes
Procedural silence	Silences audible Baby Temp alarm for 5 minutes

**SET TEMPERATURE/HEATER OUTPUT**

Set temp range	34 – 38 °C (37.0° override)
Heater output	0 – 100%, 10% increments
Skin temp display	18 – 43 °C
Accuracy	±0.2 °C (31 to 37 °C)

Not all products, features, or services are for sale in all countries.

Mentioned Trademarks are only registered in certain countries and not necessarily in the country in which this material is released. Go to [www.draeger.com/trademarks](http://www.draeger.com/trademarks) to find the current status.

**CORPORATE HEADQUARTERS**

Drägerwerk AG & Co. KGaA  
Moislinger Allee 53–55  
23558 Lübeck, Germany

[www.draeger.com](http://www.draeger.com)

**USA**

Draeger, Inc.  
3135 Quarry Road  
Telford, PA 18969-1042, USA  
Tel +1 800 4DRAGER  
(+1 800 437 2437)  
Fax +1 215 723 5935  
[info.usa@draeger.com](mailto:info.usa@draeger.com)

**Manufacturer:**

Draeger Medical Systems, Inc.  
3135 Quarry Road  
Telford, PA 18969-1042, USA

The quality management system at Draeger Medical Systems, Inc. is certified according to ISO 13485 and ISO 9001.