

## SmartCare®/PS

SmartCare®/PS is an option for the Evita XL intensive care ventilator. The SmartCare®/PS system is an automated clinical protocol, designed to stabilize the patient's spontaneous breathing in a comfortable zone of normal ventilation and to automatically reduce the ventilatory support.



### TECHNICAL DATA

SmartCare®/PS can be used for weaning intubated or tracheotomized patients with a body weight above 15 kg (33 lbs).

The weaning protocol is designed for long-term ventilated patients.

As with every weaning protocol, the patients should be ready for weaning, i.e., haemodynamically stable with adequate oxygenation and spontaneous breathing.

Patients should not have a significant V/Q mismatch e.g. pulmonary embolism. The acid-basis status should be balanced, and the patient should not have fever [1].

#### Protocol

Weaning strategy reduction	Automatic adjusted Pressure Support
Protocol implementation	Knowledge base
Metabolic parameter for classification	etCO <sub>2</sub>
Respiratory parameter for classification	f <sub>spont</sub> , VT
Data acquisition intervall	10 sec
Classification of ventilatory situation	Every 2 min/5 min
Classification limits for body weight ranges	15 kg to 35 kg (33 lbs to 77 lbs)
	35 kg to 55 kg (77 lbs to 121 lbs)
	Above 55 kg to 200 kg (121 lbs to 441 lbs)
Protocol with therapeutic measures for	Tachypnoea, Severe Tachypnea, Insufficient Ventilation, Hypoventilation, Central Hypoventilation, Hyperventilation, Unexplained Hyperventilation
Spontaneous Breathing Trial (SBT)	Automatic
Notification for readiness to separate from ventilator	Automatic
Notification for user action to change PEEP	Automatic
Combination with other options	Automatic Tube Compensation (ATC™) for patients above 35 kg (77 lbs) of body weight
Pressure change limit	Max. 4 cmH <sub>2</sub> O



MT-0617-2008

**Settings**

Medical history	COPD, Neurological Disorder
Body weight	15 kg to 200 kg (33 lbs to 441 lbs)
Airway access	Endotracheal, tracheotomized
Humidification	Active humidifier, HME
Night rest	Weaning pause, Time, Length
Manual pressure support override	At any time

**Monitoring**

Trend	Classification, SC- $P_{supp}$ Time range last 1- 24 h
Logbook	Classification, Phase, Automatic and manual changes in $P_{supp}$
Remote access	Built in Web Server

**Safety**

Alarms	Independent regular ventilator alarms Additional alarms for SmartCare®/PS
Apnea ventilation	Automatic

## Reference:

- [1] MacIntyre, N.; and the writing committee and Task force of the American College of Chest Physicians. Evidence-Based Guidelines for weaning and discontinuing Ventilatory support. Chest 2001;120:375S-395S.

**CORPORATE HEADQUARTERS**

Drägerwerk AG & Co. KGaA  
Moislinger Allee 53–55  
23558 Lübeck, Germany

[www.draeger.com](http://www.draeger.com)

**Manufacturer:**

Drägerwerk AG & Co. KGaA  
Moislinger Allee 53–55  
23558 Lübeck, Germany

**REGION EUROPE CENTRAL  
AND EUROPE NORTH**

Drägerwerk AG & Co. KGaA  
Moislinger Allee 53–55  
23558 Lübeck, Germany  
Tel +49 451 882 0  
Fax +49 451 882 2080  
[info@draeger.com](mailto:info@draeger.com)

**REGION EUROPE SOUTH**

Dräger Médical S.A.S.  
Parc de Haute Technologie  
d'Antony 2  
25, rue Georges Besse  
92182 Antony Cedex, France  
Tel +33 1 46 11 56 00  
Fax +33 1 40 96 97 20  
[dlmfr-contact@draeger.com](mailto:dlmfr-contact@draeger.com)

**REGION MIDDLE EAST, AFRICA**

Drägerwerk AG & Co. KGaA  
Branch Office  
P.O. Box 505108  
Dubai, United Arab Emirates  
Tel +971 4 4294 600  
Fax +971 4 4294 699  
[contactuae@draeger.com](mailto:contactuae@draeger.com)

**REGION ASIA / PACIFIC**

Draeger Medical  
South East Asia Pte Ltd.  
25 International Business Park  
#04-27/29 German Centre  
Singapore 609916, Singapore  
Tel +65 6572 4388  
Fax +65 6572 4399  
[asia.pacific@draeger.com](mailto:asia.pacific@draeger.com)

**REGION NORTH AMERICA**

Draeger Medical, Inc.  
3135 Quarry Road  
Telford, PA 18969-1042, USA  
Tel +1 215 721 5400  
Toll-free +1 800 437 2437  
Fax +1 215 723 5935  
[info.usa@draeger.com](mailto:info.usa@draeger.com)

**REGION CENTRAL  
AND SOUTH AMERICA**

Dräger Panama Comercial  
S. de R.L.  
Complejo Business Park,  
V tower, 10th floor  
Panama City  
Tel +507 377 9100  
Fax +507 377 9130  
[contactcsa@draeger.com](mailto:contactcsa@draeger.com)

Locate your Regional Sales  
Representative at:  
[www.draeger.com/contact](http://www.draeger.com/contact)

